



Test Certificate

OFI Technologie & Innovation GmbH
Firmensitz: 1030 Wien, Franz-Grill-Straße 5, Objekt 213
t: +43 1 798 16 01 - 0 | office@ofi.at | www.ofi.at

with reference to OFI Test Report No. 1900436/7186/9 & 7360/9 from 2019-05-07

Client: Semperit Technische Produkte GmbH
Modecenterstraße 22
1030 Vienna
AUSTRIA

Order: Test of single use gloves
for compliance with ÖNORM EN 455-1, -2, -3

Samples: Medical examination gloves Semperguard ® vinyl powder free,
VCS-048CL

Receipt of Samples: 2019-02-27, 2019-04-17


Tests: Tests were carried out in the individual technical departments of OFI within the scope of competence of the authorised signatories according to OFI QM manual.

Test Results:

Test	Test method	Date of testing	Result
Examination gloves for single use - Testing for leak tightness, (Acceptance Quality Level: AQL 1.5)	ÖNORM EN 455-1 (2001)	2019-03-05 - 2019-05-07	Compliance
Examination gloves for single use - Determination of breaking force Determination of dimensions	ÖNORM EN 455-2 (2015)	2019-03-05 - 2019-05-07	Compliance
Examination gloves for single use - Determination of residual powder	ÖNORM EN 455-3 (2015)	2019-03-05 - 2019-05-07	Compliance

The testing results relate only to the samples analysed and refer to the detailed information of OFI Test report No. 1900436/7186/9 & 7360/9.

Approved:


DI Gabriele Ettenberger
Director in charge,
Pharma, Medical Devices and Hygiene

Date:

07.05.2019

This certificate comprises 1 page.

It may be made available to third parties, either free of charge or against payment, if the full wording and if OFI is expressly named as the author. All tests applied are subject to a quality assurance programme according to EN ISO/IEC 17025:2017.